

What Are The Crystals On The Probe?

During storage, salt crystals will often form on the electrode, around the soaker bottle. Sometimes the formation is so extensive there will be crystals under the soaker bottle and on the foam inside the box. This will not affect the performance, accuracy or life of the probe. Rinse off the crystals with water prior to use.

The contents of this publication are presented for information purposes only, and while every effort has been made to ensure their accuracy, they are not to be construed as warranties or guarantees, express or implied, regarding products or services described herein or in their use or applicability. All sales are governed by our terms and conditions, which are available on request. We reserve the right to modify or improve the design or specification of such products at any time.

© 2026 Broadley-James Corporation. All rights reserved. Visit www.broadley-james.com/trademarks for trademark information.

FAQ-228
TMP-BF-102101